

<b>Sample #</b>	<b>SNP Call</b>	<b>Signal Detection</b>
DNA0001	99.10%	99.99%
DNA0002	98.32%	99.95%
DNA0004	99.15%	99.97%
DNA0005	99.02%	99.99%
DNA0008	98.81%	99.93%
DNA0009	99.66%	99.99%
DNA0010	94.43%	99.19%
DNA0011	99.19%	100.00%
DNA0012	96.61%	99.71%
DNA0013	98.60%	99.95%
DNA0014	98.58%	99.96%
DNA0015	98.92%	99.89%
DNA0016	99.06%	99.99%
DNA0017	98.74%	99.94%
DNA0018	98.85%	99.96%
DNA0019	99.37%	99.99%
DNA0020	99.63%	99.99%
DNA0021	99.55%	99.99%
DNA0022	99.49%	99.98%
DNA0023	99.45%	100.00%
DNA0024	99.55%	100.00%
DNA0025	99.43%	99.99%
<b>Mean</b>	<b>98.80%</b>	<b>99.93%</b>
SD	0.0117	0.0018
Min.	94.43%	99.19%
Max.	99.66%	100.00%
<b>Affymetrix Recommendation:</b>	<b>&gt; 95.0%</b>	<b>&gt; 98.0%</b>